



Operation

Simple to operate and easy to interpret.



Detection

The detection is fast, the result can be interpreted in 10 minutes.



Accuracy

High sensitivity and specificity.



Hangzhou Fanttest Biotech Co., Ltd.






Product background

Adenoviruses are a large group of viruses that can cause disease in the respiratory tract, eye, digestive tract, urethra and bladder. Enteric adenovirus is an important cause of infantile diarrhea, 7%~19% of infantile diarrhea is caused by it, the incidence rate is second only to rotavirus.

Intended use

This product is used for in vitro qualitative detection of adenovirus in infant stool specimens.

Operation steps and result interpretation



- Positive (+): Two purple-red bands appear. One is located in the detection area (T), and the other is located in the quality control area (C).
- Negative (-): Only a purple-red band appears in the quality control area (C). There is no purple-red band in the detection area (T).
- Invalid: There is no purple-red band in the quality control area(C).

Product information

Product name	Specimen	Format	Pack/box	Shelf life	Storage temperature	Certificate
Adenovirus Antigen Rapid Test	Feces	Cassette /Strip	25T/50T	24months	2-30°C	(€